

Aus dem Institut für Robotik und Kognitive Systeme
der Universität zu Lübeck

Direktor:
Prof. Dr.-Ing. Achim Schweikard

Safe and clinically applicable robotized Transcranial Magnetic Stimulation

Inauguraldissertation
zur
Erlangung der Doktorwürde
der Universität zu Lübeck
— Aus der Sektion Informatik —

Vorgelegt von Lars Richter
geboren in Lübeck

Lübeck 2012

Contents

Contents	xiii
1. Introduction	1
1.1. Transcranial Magnetic Stimulation	1
1.1.1. Principle of TMS	1
1.1.2. Applications of TMS: single-pulse versus repetitive stimulation . .	3
1.1.3. TMS Coils	5
1.1.4. Motor Evoked Potentials and Motor Threshold	6
1.2. State-of-the-art: Neuro-navigated TMS	9
1.2.1. Head registration & tracking	10
1.2.2. Coil tracking	11
1.3. Robotized TMS: Combining neuro-navigation with automation	12
1.3.1. Specialized setup	13
1.3.2. Industrial robot design	14
1.4. Purpose of this work	18
1.4.1. Structure of this work	20
I. Systematic analysis and evaluation of robotized TMS in practice	23
2. The importance of robotized TMS — Stability of induced electric fields	25
2.1. Principle of end-to-end accuracy	26
2.2. Realization & Data acquisition	29
2.2.1. Head motion measurements	29
2.2.2. Electric field measurements	30
2.2.3. Typical TMS scenarios	32
2.2.4. Error calculation	33
2.2.5. Statistical analysis	33
2.3. Impact of head motion on TMS	34
2.3.1. Head motion	34

2.3.2. End-to-end accuracy	34
2.4. Consequences	40
2.5. Derived requirements for robotized TMS	42
3. Evaluation of robotized TMS — the current system in practice	43
3.1. Optimal coil orientation for TMS of the lower limb	43
3.1.1. Experimental realization	44
3.1.2. Stimulation outcomes	47
3.1.3. Relevance for TMS	51
3.2. Coil-to-scalp/cortex distance	53
3.2.1. TMS recordings	54
3.2.2. Measured motor thresholds and distances	54
3.2.3. Robotized TMS for accurate coil positioning	55
3.3. Practical evaluation of robotized TMS	55
II. Safe and clinically applicable robotized TMS	59
4. Robust real-time robot/camera calibration	61
4.1. Hand-eye calibration	63
4.2. Online calibration	66
4.2.1. Basic idea of online calibration	66
4.2.2. Marker calibration	67
4.2.3. Robust real-time calibration	68
4.2.4. Translational error estimation for marker calibration	70
4.2.5. Error calculation for online calibration	72
4.2.6. Data acquisition for evaluation	72
4.3. Evaluation of online calibration	76
4.3.1. Accuracy of marker calibration	76
4.3.2. Accuracy of online calibration	77
4.4. Benefits for robotized TMS	81
5. FT-control	83
5.1. Basic principles	84
5.1.1. Sensor calibration	85
5.1.2. Gravity compensation and tool calibration	87
5.1.3. Influence of the coil's supply cable	88

5.2. Implementation of FT-control	89
5.2.1. Setup	89
5.2.2. Hand-assisted positioning	89
5.2.3. Contact pressure control	92
5.2.4. Data acquisition for evaluation of FT-control	94
5.3. Results of FT-control	98
5.3.1. Coil calibration and gravity compensation	98
5.3.2. Hand-assisted positioning	98
5.3.3. Latency of contact pressure control	99
5.4. FT-control in the context of robotized TMS	99
6. FTA-Sensor — combination of force/torque and acceleration	101
6.1. The FTA sensor	102
6.1.1. Combining acceleration with force-torque	102
6.1.2. Embedded System for real-time monitoring	103
6.1.3. Hardware design — circuit board and casing	104
6.1.4. Calibration of IMU to FT sensor	106
6.1.5. Data acquisition for evaluation of the FTA sensor	109
6.2. Performance of the FTA sensor	112
6.2.1. Calibration	112
6.2.2. Gravity compensation	115
6.2.3. Latency	116
6.2.4. Realistic worst-case estimate	117
6.3. FTA sensor for safe robotized TMS	119
7. Optimized FT-control with FTA sensor	121
7.1. Advanced hand-assisted positioning	121
7.2. Integration into the robot server	123
7.3. TMS coil calibration	126
7.4. Data acquisition for realistic evaluation of optimized FT-control	127
7.4.1. Coil calibration and gravity compensation	127
7.4.2. Precision of optimized hand-assisted positioning	128
7.5. Performance of the FTA sensor in operation	130
7.5.1. Coil calibration and gravity compensation	130
7.5.2. Precision of optimized hand-assisted positioning	131
7.6. Optimized FT-control for clinical acceptance	133

8. Direct head tracking	135
8.1. Direct versus indirect tracking	135
8.2. FaceAPI	136
8.2.1. The FaceAPI's main principle	136
8.2.2. Evaluation of the FaceAPI for direct head tracking	136
8.2.3. Accuracy of the FaceAPI	137
8.3. 3D laser scans	137
8.3.1. Implementation of direct head tracking with laser scans	137
8.3.2. Data acquisition for an experimental validation	143
8.3.3. First results	144
8.4. Head contour generation based on laser scans	148
8.4.1. Head scanning and contour generation	149
8.4.2. Comparison to manual contour generation	150
8.4.3. Application in robotized TMS studies	151
8.5. Capability of direct tracking for robotized TMS	151
III. Discussion and closing remarks	153
9. Discussion	155
9.1. Robust real-time robot/world calibration	156
9.2. Hand-assisted positioning	158
9.3. Contact pressure control	159
9.4. FTA sensor	160
9.5. Direct head tracking	163
9.5.1. FaceAPI	163
9.5.2. 3D laser scans	164
10. Closing remarks	167
10.1. Conclusions	167
10.2. Outlook and future work	168
10.2.1. Fully automated TMS	168
10.2.2. Mapping of the spinal roots	169
10.2.3. Direct head tracking	170
10.2.4. Double-coil robotized TMS	172
10.2.5. Robotized interleaved TMS/fMRI	173
Bibliography	177

List of Figures	189
List of Tables	193
Glossary and List of Abbreviations	195
Nomenclature	201